

The manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe, hereby declares that the machinery described below:

Product: MCR Modular Compact Rheometer

Type: MCR 502, MCR 702 TwinDrive™

Serial number:

complies with the provisions of the **Machinery Directive (2006/42/EC)** and the regulations transposing it into national law

complies with the provisions of the **Electromagnetic Compatibility Directive (2004/108/EC)**

complies with the safety objectives of the **Low Voltage Directive (2006/95/EC)**

complies with the provisions of the following harmonized standards:

- EN ISO 12100:2010 Safety of machinery - General principles for design, risk assessment and risk reduction –
- EN 61326-1:2006 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

The product is classified as a class A equipment and is not intended for the use in industrial area

complies with the provisions of the following technical standards:

- IEC 61010-1 3rd ed.:2010 Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements

The manufacturer compiles the technical file according to 2006/42/EC Annex II

Done at Graz, 2013-12-10


Ing. Peter Kettisch
Division Manager


DI Dr. Wolfgang Baumgartner
Department Manager