



The Manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe hereby declares that the product listed below

Product designation: **C-LTD180/QC TEMPERATURE CONTROL SYSTEM WITH PT 100**
Model: **C-LTD180/QC**
Material number: **18658**

is in conformity with the relevant European Union harmonisation legislation.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

- **Electromagnetic Compatibility (2014/30/EU, OJ L 96/79 of 29.3.2014)**

Applied standards:

EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
-----------------	--

The product is classified as a class B equipment and is not intended for the use in industrial area.

- **Low Voltage Directive (2014/35/EU, OJ L 96/357 of 29.3.2014)**


Applied standards:

EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements
-----------------	---

EN 61010-2-010:2014	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of materials
---------------------	---

Place and date of issue: Graz, 2016-06-23


DI Günter Hofer
Executive Director
Business Unit Measurement


MSc Fatemeh Rezaei
Head of Viscometry
Business Unit Measurement