



The manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe, hereby declares that the machinery described below:

Description: High Pressure Asher
Type: HPA-S (Drawing Number B25MZ21)
Serial Number:

complies with the provisions of the **Machinery Directive (2006/42/EC)** and the regulations transposing it into national law,
(Concerning to Annex I, Section 1.2.4.3 for the Emergency Stop an alternative safety concept has been applied.)

complies with the provisions of the **Pressure Equipment Directive (97/23/EC)** and the following standards

- AD Guidelines AD2000

The pressure device has been taken a conformity assessment procedure regarding Modul B1 + F. Tests and inspections has been observed from TÜV Austria Services GmbH with the ID-No. 0408. Following documents are available:

- Inspection of the Draft: 2012-FRW-191 Rev.0 vom 27.11.2012
- Type Examination Certificate, No: 2012-HAST-0444
- Certificate of Conformity of the pressure vessel: KOB.....

complies with the provisions of the **Electromagnetic Compatibility Directive (2004/108/EC)**,

complies with the safety objectives of the **Low Voltage Directive (2006/95/EC)**,

complies with the provisions of the following harmonized standards:

- EN ISO 12100:2010 Safety of machinery – General principles for design – Risk assessment and Risk reduction
- EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

The product is classified as a class B equipment and is not intended for the use in industrial area.

complies with the provisions of the following technical standards:

- EN 61010-1:2001 Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements
- EN 61010-2-010:2003 Part 2-010: Particular requirements for laboratory equipment for the heating of materials

Ing. Peter Kettisch
Business Unit Manager

Dr. Hans Hahn
Product Line Manager