

Anton Paar GmbH hereby declares that the product listed below in the version offered for sale meets all the basic requirements of the applicable sections of the relevant EU directives in design and type.

This declaration will be deemed invalid should any unauthorized modifications be made to the product. Follow the information given in the instruction manual when setting up and operating the instrument.

Product designation: **Particle Analyzer**
Model: **Litesizer 500**
Manufacturer: **Anton Paar GmbH**

The product meets the requirements of the following directives:

- **Electromagnetic Compatibility 2004/108/EC**

Applied standards:

EN 61326-1:2013

Electrical equipment for measurement, control and laboratory use -
EMC requirements - Part 1: General requirements

The product is classified as a Class B equipment and is not intended for the use in industrial area.

- **Low Voltage Directive 2006/95/EC**

Applied standards:

EN 61010-1:2010


Safety requirements for electrical equipment for measurement,
control and laboratory use Part 1: General requirements

EN 61010-2-010:2014

Safety requirements for electrical equipment for measurement,
control, and laboratory use – Part 2-010: Particular requirements for
laboratory equipment for the heating of materials

EN 62233:2008

Measurement methods for electromagnetic fields of household
appliances and similar apparatus with regard to human exposure



DI(FH) Dr. Jakob Santner
Executive Director
Business Unit Characterization



DI Dr. Wolfgang Baumgartner
Head of Rheometry
Business Unit Characterization