

EU Declaration of Conformity

(original)



The Manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe hereby declares that the product listed below

Product designation: **Pico 3000 Compact Controller**
Model: **Pico 3000 HMI L-Dens Ex d set**
Pico 3000 RC Housing wall mounting Ex d
Pico 3000 RC Housing wall mounting NPT Ex d
Material number: 177598, 179380, 191812,

is in conformity with the relevant European Union harmonisation legislation. This declaration of conformity is issued under the sole responsibility of the manufacturer.

Electromagnetic Compatibility (2014/30/EU, OJ L 96/79 of 29.3.2014)

Applied standards:

- EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
- EN 61326-2-3:2013 Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-3: Particular requirements – Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

The product is classified as a class A equipment and is intended for the use in industrial area.

ATEX Directive (2014/34/EU, OJ L 96/309 of 29.3.2014)

Applied standards:

- IEC 60079-0:2017, IEC 60079-1:2014, EN 60079-0:2012, EN 60079-0:2012/A1:2013

Notified body: Identification number: 0123, TÜV SÜD Product Service GmbH / Germany / 80339 München, Ridlerstraße 65 / EU-Type Examination Certificate Number: TPS 18 ATEX 18013 013 X


Safety objectives of the Low Voltage Directive (2014/35/EU, OJ L 96/357 of 29.3.2014)


Applied standards:

- EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements
- EN 61010-2-201:2013 Safety requirements for electrical equipment for measurement, control and laboratory use - Part 2-201: Particular requirements for control equipment

RoHS Directive (2011/65/EU, OJ L 174/88 of 1.7.2011)

Place and date of issue: Graz, 2018-04-11


ppa P. Kettisch
Ing. Peter Kettisch
Executive Director
Business Unit Solutions


Mag.(FH), BSc Rainer Pirchegger
Head of Process Instrumentation
Business Unit Solutions