

# EU Declaration of Conformity

(original)



The Manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe hereby declares that the product listed below

Product designation: **OXY 4100 TRANSMITTER HMI / TRACE RANGE**  
**OXY 4100 TRANSMITTER HMI / WIDE RANGE**

Model: **Oxy 4100 Transmitter Trace Range**  
**Oxy 4100 Transmitter Wide Range**

Material number: 261588, 261589

is in conformity with the relevant European Union harmonisation legislation. This declaration of conformity is issued under the sole responsibility of the manufacturer.

## Electromagnetic Compatibility (2014/30/EU, OJ L 96/79 of 29.3.2014)

Applied standards:

- EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements
- EN 61326-2-3:2013 Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-3: Particular requirements – Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning


## Low Voltage Directive (2014/35/EU, OJ L 96/357 of 29.3.2014)

Applied standards:


- EN 61010-1:2010 +1:2019 +A1:2019/AC:2019 Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements

## RoHS Directive (2011/65/EU, OJ L 174/88 of 1.7.2011)

Place and date of issue: Graz, 2023-01-18

DocuSigned by:  
  
 4C1800E6A5304C2...

DI Dr. Christopher Fradler, MBA  
 Executive Director  
 Business Unit Solution

DocuSigned by:  
  
 34D9DC346C704C5...

Mag.(FH) Rainer Pirchegger, BSc  
 Head of Process Instrumentation  
 Business Unit Solution