

UK Declaration of Conformity



The Manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe hereby declares that the product listed below

Product designation: **OXY 4100 TRANSMITTER HMI / TRACE RANGE**
OXY 4100 TRANSMITTER HMI / WIDE RANGE

Model: **Oxy 4100 Transmitter Trace Range**
Oxy 4100 Transmitter Wide Range

Material number: 261588, 261589

is in conformity with all the relevant UK legislation

Electrical Equipment (Safety) Regulations 2016, 2016 No. 1101

Electromagnetic Compatibility Regulations 2016, 2016 No. 1091

Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012, 2012 No. 3032


complies with the designated standards:

- EN 61010-1:2010 +A1:2019 +A1:2019/AC:2019
- EN 61326-1:2013


This declaration of conformity is issued under the sole responsibility of the manufacturer.

Importer: Anton Paar Ltd, Unit F, The Courtyard, Hatfield Rd, St Albans AL4 OLA, United Kingdom;

Place and date of issue: Graz, 2023-01-18

DocuSigned by:

4C1800E6A5304C2...

DI Dr. Christopher Fradler, MBA
Executive Director
Business Unit Solution

DocuSigned by:

34D9DC346C704C5...

Mag.(FH) Rainer Pirchegger, BSc
Head of Prozess Instrumentation
Business Unit Solution