

# EU Declaration of Conformity

(original)



The Manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe hereby declares that the product listed below

Produktbezeichnung: **L-Cor xx00 mass flow meter**

Typenbezeichnung Sensor: **AUxxx - Lxxxx - xxx - xxxxxB – 0**  
**ABxxx - Lxxxx - xxx - xxxxxB – 0**  
**ASxxx - Lxxxx - xxx - xxxxxB – 0**

Typenbezeichnung Transmitter: **PA0K xxxxx**

Materialnummer: **800000 - 810000**

is in conformity with the relevant European Union harmonisation legislation. This declaration of conformity is issued under the sole responsibility of the manufacturer.

## Electromagnetic Compatibility (2014/30/EU, OJ L 96/79 of 29.3.2014)

Applied standards:

- EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

## ATEX Directive (2014/34/EU, OJ L 96/309 of 29.3.2014)

Applied standards: EN IEC 60079-0:2018, EN 60079-1:2014, EN 60079-11:2012

EU – Type Examination Certificate Number: DEKRA 23ATEX0093 X

Marking:



II 2G Ex db ib IIC T4 Gb  
 II 2G Ex db ib IIB T4 Gb  
 II 2G Ex db [ib] IIC T6 Gb

EU – Type Examination Certificate Number: DEKRA 23ATEX0094

Marking:



II 2G Ex ib IIC T1 ...T5 Gb  
 II 2G Ex ib IIB T3 ...T4 Gb

Certification Body: DEKRA Certification B. V., Identification number: 0344, Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 8602 ED Arnhem, Netherlands

Notified Body: TÜV AUSTRIA SERVICES GMBH, Identification number: 0408, Deutschstrasse 10, 1230 Wien, Austria

## Safety objectives of the Low Voltage Directive (2014/35/EU, OJ L 96/357 of 29.3.2014)

Applied standards:

- EN 61010-1:2010 + A1:2019 Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements  
 + A1:2019/AC:2019

## RoHS Directive (2011/65/EU, OJ L 174/88 of 1.7.2011 + 2015/863/EU)

Place and date of issue: Graz, 2023-12-15

DocuSigned by:  
  
 4C1800E6A5304C2

DI Dr. Christopher Fradler, MBA  
 Executive Director  
 Business Unit Solution

DocuSigned by:  
  
 C4B0FF0B8EFF4AD...

Dr. Christoph Ebner  
 Head of Process Instrumentation  
 Business Unit Solutions