

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: **781 762 4300** F: 781-762-9375 www.fmapprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

This certificate is issued for the following equipment:

135165, L-Vis 520 Ex Smart Sensor, Inline Viscometer

Equipment Ratings:

Explosionproof for Class I, Division 1, Groups C and D hazardous locations, T6 Ta -20°C to +40°C with dual seal rating for maximum pressure rating of 25 Bar, process temperature up to 195°C and maximum viscosity of 50,000 mPa.s. The ingress protection rating is IP65.

FM Approved for:

Anton Paar GmbH
Graz, Austria

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

CSA-C22.2 No. 0.4	R2013
CSA-C22.2 No. 0.5	R2012
CSA-C22.2 No. 30	R2012
CAN/CSA-C22.2 No. 61010-1	2012
CAN/CSA C22.2 No. 60529	R2010
ANSI/ISA-12.27.01	2011

Original Project ID: 3049003C

Approval Granted: October 3, 2014

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC



J.E. Marquedant
Manager, Electrical Systems

3 October 2014

Date