## **UK Declaration of Conformity**



The Manufacturer **Anton Paar GmbH**, Anton-Paar-Str. 20, A-8054 Graz, Austria – Europe hereby declares that the product listed below

Product designation: SVM 2001

**SVM 4001** 

Model: SVM<sup>™</sup> 2001 Stabinger Viscometer<sup>™</sup>

SVM<sup>™</sup> 4001 Stabinger Viscometer<sup>™</sup>

Material number: 105000, 105002

is in conformity with all the relevant UK legislation

Electrical Equipment (Safety) Regulations 2016, 2016 No. 1101

Electromagnetic Compatibility Regulations 2016, 2016 No. 1091

Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012, 2012 No. 3032

complies with the designated standards:

- EN 61010-1:2010 + A1:2019 + A1:2019/AC:2019
- EN 61010-2-010:2014
- EN 61326-1:2013

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Importer: Anton Paar Ltd, Unit F, The Courtyard, Hatfield Rd, St Albans AL4 OLA, United Kingdom;

Place and date of issue: Graz, 2022-04-20

17130DDD5260426.

DI Steffen Riemer, MBA Executive Director

Business Unit Measurement

Fatemeh Rezai, MSc. Head of Viscometry Business Unit Measurement